



TEST REPORT

EN 14563, Tuberculocidal Activity

Client Gama Health Ltd

Date

04 12 2011

Product

Clinell Universal

Appearance : Clear liquid after extraction by pressure

Conditions

Temperature 20'c

Dirty conditions

Contact time 60 mins

Diluent Hard water

Neutraliser 5% VV Egg Yolk : 40g/l Polysorbate 80

.....  
D.O'Connor B.Sc. Ci.Biol M.I.F.S.T.

Validation Controls	Test Organism	Mycobacterium terrae 15755	
Validation Suspension	Met required criteria		
Control A	Met required criteria		
Control B	Met required criteria		
Control C	Met required criteria		
Test suspension	Met required criteria	9.46	log cfu/ml
Water control	Met required criteria	8.94	log cfu/ml

Concentration of				
Pass dilution	Na	log Na	Log Reduction	Contact Time
neat 100%	104	2.02	6.92	60 mins 20°C

Log reduction under dirty conditions 6.92

The test substance > Log 4 reduction which meets the specified criteria.